

**Recommended Spare Parts**
**SP2000-H/DIL-BR 115V Gas Sample Probe**

<b>Manufacturer P/N</b>	<b>Cemtrex P/N</b>	<b>Description</b>
90S0020	20-336-2000	Filter element type S-2K150. Length: 150mm, material: ceramics, filter porosity: 2µm.
90F0125	20-307-2314	Filter element type F-0,1GF150. Length: 150mm, material: glass fiber, filter porosity: 0,1µm.
20S4300	20-336-1008	Complete set SP2000H/DIL critical orifice incl. 2x O-ring. Material: glass # please indicate ratio with the order.
93S2083	20-307-2315	Special spun-glass hightemperature resistant for probes with /FW insert. Contents:1000g.
93S2096	20-307-2316	Spun glass cartridge element FW with Demister, stainless steel, without filling.
93S0045	20-336-5000	Gasket (30) for filter element. Material: Viton.
93S0051	20-307-2318	Adaptor for probe filter element F-0,1GF150 Material PTFE.
93S0020	20-336-3000	O-ring (39) for SP2000 and CLF. Material: Viton.
93S0025	20-336-4000	O-ring (55) for SP2000. Material: Viton.
93S0035	20-30-2319	Sealing spiral o-ring (39) for SP2000/7aT. Material: PTFE.
93S0040	20-307-2320	Sealing spiral o-ring (55) for SP2000/7aT. Material: PTFE.
90S2077	20-307-2322	Flange gasket DN65 PN6B (67). Material: Novapress.
93S2350	20-336-6000	Flange gasket 3" ANSI.
90S2075	20-307-2324	Flange gasket set for DN65 PN6B consisting of gasket (67) and a set of screws M12x60
93S0010	20-307-2338	Thermostat (0-180°C), with over-temperature limiter and low-temperature alarm for probe series SP.
93S0018	20-307-2340	Temperature resistant heat sink compound, 100g-tube.
93S0015	20-307-2341	Heating cartridge for SP2000-H, L=160mm, 230V/800W.
93S0017	20-336-1003	Heating cartridge for SP2000-H, L=160mm, 115VAC/800W.
93S0059	20-307-2342	PT100 sensor for SP2000H,,SP3200
93S0060	20-307-2385	Spare thermoelement Fe-CuNi for SP.../H320.
93S0061	20-336-1012	Spare thermocouple NiCr-Ni for SP.
93S0019	20-307-2344	Aluminium body for heating cartridge SP2000H.
93S0085	20-307-2386	Press-Screw M8 for filter housing lid SP2000/2100-H.
93S2125	20-307-2326	Clamp LK 115 for filter housing lid SP2000H.
93S0065	20-307-2345	Mounting clip 1 1/4" for heated sample line and SP2000-H.

<b>Manufacturer P/N</b>	<b>Centrex P/N</b>	<b>Description</b>
93S2085	20-307-2346	Filter housing lid compl. for SP2000-H with viton-o-ring,filterholder and press screw up to Ser.no.10282
93S2087	20-307-2327	Filter housig lid compl for SP2000./7aT with PTFE gasket lid seal filter element clamp. Up to SNo.10283
93S2080	20-307-2328	Filter housing lid compl. for SP2000-H/FW. Spun glass filter element without filling. With Viton o-rings and flat gasket made of Novapress up to SNo. 10283
93S2084	20-307-2329	Filter screw M6 for SP2000, 1.4571.
93S0090	20-307-2349	Hand screw M6 for SP2000 connection protection.
93S0100	20-336-1010	Blowback/calibration gas valve for probe type SP2XXX and SP3XXX with integrated filter, Working pressure: 0,7bar, Connection: 1/4"NPT-DN4/6 Material: SS316, Viton.
93S0070	20-307-2350	Spare weather protection shield for SP2000-H and SP2000-H/DIL.
93S2353	20-307-2426	Flange gasket 3" ANSI. with hole for BR Material Novapress
93S4006	20-307-2427	Dilution cross 180°C/320°C empty, without any inserts Material SS316
93S4016	20-307-2430	Bypass T-piece 180°C/320°C empty, without any inserts Material SS316
93S4009	20-307-2428	Gasket set for critical orifice of probe SP2000-H/DIL and SP2006-H/DIL, 2x Viton O-Ring 2,9 x 1,78.