

Recommended Spare Parts
SP2000-H/DIL-B 230V Gas Sample Probe

| Manufacturer P/N | Cemtrex P/N | Description |
|-------------------------|--------------------|--|
| 90S0020 | 20-336-2000 | Filter element type S-2K150. Length: 150mm, material: ceramics, filter porosity: 2µm. |
| 90F0125 | 20-307-2314 | Filter element type F-0,1GF150. Length: 150mm, material: glass fiber, filter porosity: 0,1µm. |
| 20S4300 | 20-336-1008 | Complete set SP2000H/DIL critical orifice incl. 2x O-ring. Material: glass # please indicate ratio with the order. |
| 93S2083 | 20-307-2315 | Special spun-glass hightemperature resistant for probes with /FW insert. Contents:1000g. |
| 93S2096 | 20-307-2316 | Spun glass cartridge element FW with Demister, stainless steel, without filling. |
| 93S0045 | 20-336-5000 | Gasket (30) for filter element. Material: Viton. |
| 93S0051 | 20-307-2318 | Adaptor for probe filter element F-0,1GF150 Material PTFE. |
| 93S0020 | 20-336-3000 | O-ring (39) for SP2000 and CLF. Material: Viton. |
| 93S0025 | 20-336-4000 | O-ring (55) for SP2000. Material: Viton. |
| 93S0035 | 20-30-2319 | Sealing spiral o-ring (39) for SP2000/7aT. Material: PTFE. |
| 93S0040 | 20-307-2320 | Sealing spiral o-ring (55) for SP2000/7aT. Material: PTFE. |
| 90S2077 | 20-307-2322 | Flange gasket DN65 PN6B (67). Material: Novapress. |
| 93S2350 | 20-336-6000 | Flange gasket 3" ANSI. |
| 90S2075 | 20-307-2324 | Flange gasket set for DN65 PN6B consisting of gasket (67) and a set of screws M12x60 |
| 93S0010 | 20-307-2338 | Thermostat (0-180°C), with over-temperature limiter and low-temperature alarm for probe series SP. |
| 93S0018 | 20-307-2340 | Temperature resistant heat sink compound, 100g-tube. |
| 93S0015 | 20-307-2341 | Heating cartridge for SP2000-H, L=160mm, 230V/800W. |
| 93S0017 | 20-336-1003 | Heating cartridge for SP2000-H, L=160mm, 115VAC/800W. |
| 93S0059 | 20-307-2342 | PT100 sensor for SP2000H,,SP3200 |
| 93S0060 | 20-307-2385 | Spare thermoelement Fe-CuNi for SP.../H320. |
| 93S0061 | 20-336-1012 | Spare thermocouple NiCr-Ni for SP. |
| 93S0019 | 20-307-2344 | Aluminium body for heating cartridge SP2000H. |
| 93S0085 | 20-307-2386 | Press-Screw M8 for filter housing lid SP2000/2100-H. |
| 93S2125 | 20-307-2326 | Clamp LK 115 for filter housing lid SP2000H. |
| 93S0065 | 20-307-2345 | Mounting clip 1 1/4" for heated sample line and SP2000-H. |

| Manufacturer P/N | Centrex P/N | Description |
|-------------------------|--------------------|---|
| 93S2085 | 20-307-2346 | Filter housing lid compl. for SP2000-H with viton-o-ring,filterholder and press screw up to Ser.no.10282 |
| 93S2087 | 20-307-2327 | Filter housig lid compl for SP2000./7aT with PTFE gasket lid seal filter element clamp. Up to SNo.10283 |
| 93S2080 | 20-307-2328 | Filter housing lid compl. for SP2000-H/FW. Spun glass filter element without filling. With Viton o-rings and flat gasket made of Novapress up to SNo. 10283 |
| 93S2084 | 20-307-2329 | Filter screw M6 for SP2000, 1.4571. |
| 93S0090 | 20-307-2349 | Hand screw M6 for SP2000 connection protection. |
| 93S0100 | 20-336-1010 | Blowback/calibration gas valve for probe type SP2XXX and SP3XXX with integrated filter, Working pressure: 0,7bar, Connection: 1/4"NPT-DN4/6 Material: SS316, Viton. |
| 93S0070 | 20-307-2350 | Spare weather protection shield for SP2000-H and SP2000-H/DIL. |
| 93S2353 | 20-307-2426 | Flange gasket 3" ANSI. with hole for BR Material Novapress |
| 93S4006 | 20-307-2427 | Dilution cross 180°C/320°C empty, without any inserts Material SS316 |
| | | |
| 93S4009 | 20-307-2428 | Gasket set for critical orifice of probe SP2000-H/DIL and SP2006-H/DIL, 2x Viton O-Ring 2,9 x 1,78. |