

**Recommended Spare Parts**
**SP2006-H/DIL-BR/2x/FRP Gas Sample Probe**

| <b>Manufacturer P/N</b> | <b>Cemtrex P/N</b> | <b>Description</b>   |
|-------------------------|--------------------|--|
| 90S0020                 | 20-336-2000        | Filter element type S-2K150. Length: 150mm, material: ceramics, filter porosity: 2µm.              |
| 90F0125                 | 20-307-2314        | Filter element type F-0,1GF150. Length: 150mm, material: glass fiber, filter porosity: 0,1µm.      |
| 93S0045                 | 20-336-5000        | Gasket (30) for filter element. Material: Viton.   |
| 93S0051                 | 20-307-2318        | Adaptor for probe filter element F-0,1GF150 Material PTFE.   |
| 93S0020                 | 20-336-3000        | O-ring (39) for SP2000 and CLF. Material: Viton.   |
| 93S0025                 | 20-336-4000        | O-ring (55) for SP2000. Material: Viton.   |
| 90S2077                 | 20-307-2322        | Flange gasket DN65 PN6B (67). Material: Novapress.   |
| 93S2350                 | 20-336-6000        | Flange gasket 3" ANSI.   |
| 93S0010                 | 20-307-2338        | Thermostat (0-180°C), with over-temperature limiter and low-temperature alarm for probe series SP. |
| 93S0015                 | 20-307-2341        | Heating cartridge for SP2000-H, L=160mm, 230V/800W.  |
| 93S0017                 | 20-336-1003        | Heating cartridge for SP2000-H, L=160mm, 115VAC/800W.  |
| 93S0018                 | 20-307-2340        | Temperature resistant heat sink compound, 100g-tube.   |
| 93S0059                 | 20-307-2342        | PT100 sensor for SP2000H,,SP3200   |
| 93S0061                 | 20-336-1012        | Spare thermocouple NiCr-Ni for SP.   |
| 93S0083                 | 20-307-2335        | Clamp clip NEW 04/04 SP2000.   |
| 93S0085                 | 20-307-2386        | Press-Screw M8 for filter housing lid SP2000/2100-H.   |
| 93S2125                 | 20-307-2326        | Clamp LK 115 for filter housing lid SP2000H.   |
| 93S0065                 | 20-307-2345        | Mounting clip 1 1/4" for heated sample line and SP2000-H.  |
| 93S1985                 | 20-336-1016        | Filter housing lid compl. with viton-o-ring,filterholder and press screw. From S-No. 10283         |
| 93S2084                 | 20-307-2329        | Filter screw M6 for SP2000, 1.4571.  |
| 93S0090                 | 20-307-2349        | Hand screw M6 for SP2000 connection protection.  |
| 93S0070                 | 20-307-2350        | Spare weather protection shield for SP2000-H and SP2000-H/DIL.                                     |
| 93S4425                 | 20-336-9000        | Bypass-unit complete for SP2006-H/DIL-BR.  |
| 93S4401                 | 20-336-1015        | Injector complete unit SP2006-H/DIL incl. O-rings and sample outlet tube                           |
| 93S4406                 | 20-336-1013        | Dilution unit for SP2006-H/DIL 180°C/320°C empty, without any inserts Material SS316.              |

| <b>Manufacturer P/N</b> | <b>Centrex P/N</b> | <b>Description</b>   |
|-------------------------|--------------------|--|
| 93S4411                 | 20-307-2432        | Injector complete unit SP2006-H/B incl. O-rings and sample outlet tube                     |
| 93S4416                 | 20-336-1014        | Bypass unit SP2006H/DIL 180°C/320°C empty, without any inserts Material SS316              |
| 93S4420                 | 20-336-8000        | Dilution unit complete for SP2006-H/DIL. DIL-ratio to be specified with the order.         |
| 93S4412                 | 20-336-1004        | Gasket set for DIL unit of probe SP2006-H/DIL, Viton.                                      |
| 93S4413                 | 20-336-1005        | Gasket set for Bypass unit of probe SP2006-H/DIL, Viton.                                   |
| 93S4415                 | 20-307-1751        | Gasket set complete for probe SP2006-H/320/DIL/BR incl. top lid sealings Chemraz 615 16 ea |
| 93S4435                 | 20-336-1009        | Sucktube SP2006H/DIL incorporated filter, without O-rings Material: SS316.                 |
| 93S4404                 | 20-307-2433        | Complete unit SP2006-H/DIL incl. O-rings and sample outlet tube                            |